

website : www.medical-developpement.com

Training Treadmill : S1830

Product datasheet



MEDICAL DEVELOPPEMENT

SARL au capital de 100 000€ – SIRET 799 227 152 00011 – NAF 3230Z – N° TVA : FR63 799 227 152

17 ZA du pré de l'orme – 38760 VARCES ALLIERES ET RISSET

Main presentation

The new large size generation of training treadmill, like the **S1830** Treadmill is especially designed and dimensioned to allow medical and research activities in the fields like running, roller-skating, roller-skiing, ... who need a large surface area.

- The maximum normal speed is 30km/h (possibility to increase until 40km/h)
- Slope adjustment from -8% to +25%
- Adjustable lateral and front guard rails

For the user security, the use of an harness, especially for speed over 10km/h is required.

Area and limit of use :

This device is designed for the use below :

- **Walking and running** : 2 users can use it together
- **Roller-skating** : 1 user
- **Roller-skiing** : 1 user
- **Wheelchair** : 1 user
- **Bicycle**

To prevent the warming during high speed bicycling, the support of the sole is equipped, on the right side of the strip, a closed loop water cooling system.

Strip : The treadmill is equipped with a special thick strip, which can accept mechanical stress, especially for roller-skiing